



Guided Surgery Tool Kit (GSTK) & Master Sleeves

Catalog & Manual



Plan your case using Alpha-Bio Tec.'s implants and your Guided Surgery Planning Software of choice



The Kit

The kit comprises a variety of drills and tools required for performing the entire guided surgery procedure.

- Ergonomically and carefully designed autoclavable box and tray
- Compatible master sleeves
- Materials:
 - O Box and tray Radel® polyphenylsulfone resin
 - O Tools and drills stainless steel
 - O Sleeves stainless steel







Use the software of your choice with Alpha-Bio Tec. Kit!

















The Tray

The tray features a modular layout. Its content is organized to support the entire guided surgery procedure from site preparation to final implantation.

All kit components fit the matching master sleeves.





SITE PREPARATION

Contains the tissue punch and crestal drills required for preparing the site, and the drills and pins required for anchoring the guided osteotomy.



OSTEOTOMY

Contains the color-coded drills required by our drilling sequence for performing the guided osteotomy.



IMPLANT PLACEMENT

Contains implant mounters and their associated screws required for placing the implant through the guide. The implant mounter is attached to the implant using the associated screws, allowing the implant to be placed through the guide for increased accuracy and predictability of the implant position.






TOOLS & ACCESSORIES

Contains a variety of tools and accessories required for performing the surgery procedure without the need of supplementary external tools.





Master Sleeves


Two different sleeve diameters are available depending on the implant diameter used:

| | | Sleeve Adaptor | | |
|----------|---|--|-------------------------------------|--|
| | Used for drilling an | Used for drilling through the large sleeve (SLL) | | |
| | Ø4.1 mm | Ø5.5 mm | 11 | Ø4.1 mm |
| | 103 | 103 | Ü | B |
| Code | SLS | SLL | SLSE | SLSA |
| Ref. no. | 66012 | 66013 | 66014 | 65058 |
| Info. | For Ø 3.2, Ø 3.3, Ø 3.5, Ø 3.7N, Ø 3.75 implants | For Ø 4.2, Ø 4.65, Ø 5 and Ø 5.3 implants | For 1.5 mm drill and lateral pin | Must be removed befor implant insertion. See note below. |

Note

When using the \emptyset 5.5 mm sleeve, an adaptor should be used to mitigate the tool size. If sleeve adaptors are used for the site preparation and osteotomy stages, they **must** be removed **before** inserting the implant through the guide (when applicable).

Drills and implant mounters are prolonged by fix 9 mm to meet the extra height attained by the surgical guide, i.e. the tools' stopper is located **exactly 9 mm above** the implant level.





Modular Content

The kit is available in two different configurations or a full kit option, allowing users to choose the content according to their clinical needs. In addition, each component has its own reference number and can be purchased separately.




Entry-Level Option (#KIT65080)

Parts enable fixation of the guide and site preparation for the most popular implant sizes.

Content:

- ✓ Tray and Box
- ✓ Tissue punch (2)
- ✓ Crestal Drill (4)
- ✓ Lateral Drill & Pin (4)
- ✓ Drills (29)
- ✓ Sleeve Adaptor (2)
- Stainless steel bath with company logo
- ✓ Large autoclaveable box with company logo on top



Entry-Level Extension (#KIT65081)


Extend surgery options with additional parts, which also enable placement of implants of the most popular diameters.

Content:

- ✓ Implant placement mounters (16)
- ✓ Screws (16)
- ✓ Accessories
- ✓ Drivers (5)
- ✓ Adapters (4)
- ✓ Implant Extractor (1)

Note

This option does not include the box and tray.



Full Kit (#KIT65000)

Enables full surgery performance, from fixation of the guide to implant placement for all Apha-Bio Tec. implant diameters.

Content:

- ✓ Tray and box
- ✓ Tissue punch (2)
- ✓ Crestal Drill (4)
- ✓ Lateral Drill & Pin (4)
- ✓ Drills (40)
- ✓ Sleeve Adaptor (2)
- ✓ Implant Placement Mounters (16)
- ✓ Screws (16)
- ✓ Accessories

- ✓ Drivers (5)
- ✓ Adapters (4)
- ✓ Implant Extractor (1)
- ✓ Stainless steel
- bath with company logo

 ✓ Large autoclaveable
- box with company logo on top





To achieve the best clinical results and accuracy with Alpha-Bio $_{\text{Tec}}$'s Guided Surgery Tool Kit, perform the surgery and follow our recommended drill sequence.



Preparing for Surgery

Cold Sterilization Prior to surgery, the surgical guide must be sterilized according to the manufacturer's instructions and method used. The solution must not contain more than 15% alcohol and the process should not exceed 30 minutes. The guide must then be washed with sterile saline solution before checking it in the patient's mouth.

⚠ The guide should not be heat sterilized! Keep the guide in a cool place and avoid direct exposure to heat and humidity.

Proper Fit of the Surgical Guide The accuracy of the surgical guide requires proper fit and positioning in the patient's mouth.

⚠ The guide must sit securely without any "rocking". Checking the fit of the guide in the mouth prior to the surgery date is advised.

2




Tissue Punch

Position the surgical guide in place, hold it and drive the tissue punch through the guide. Then remove the guide and manually complete the alveolar tissue clearing.

Note

Securing the guide at this stage is not required but is recommended in the case of edentulous patients.





In order to avoid intraoperative displacements of the surgical guide after repositioning (that may affect the accuracy of the surgery), secure the guide in place.

There are two ways to achieve this:




Using 2-3 lateral pins (edentulous cases) to hold the guide between the two arches during fixation.



Using crestal pins with the use of the dental osteotomies (in full or partial edentulous cases).


Note

Fully securing the teeth supported guide is not required in most cases; however, since the guide is NOT self-retentive, it should be held in position at every stage.



Crestal Drill

While using the crestal drill, remove any interference from the alveolar crest contour. Drill all the way, until the stopper is engaged with the guide sleeve.




Pilot Drill

Start the drilling sequence using the D2.0x8 mm pilot drill. To enable drilling through the large sleeves, use the corresponding sleeve adapter.

Drill all the way, until the stopper is engaged with the guide sleeve.









Drilling Sequence

Continue with the drilling sequence, according to the implant diameter and length.* To enable drilling through the large sleeves, use the corresponding sleeve adapter. Drill all the way, until the stopper is engaged with the guide sleeve.

Note

If sleeve adaptors were used for drilling, remove them first, before continuing with implant placement.

Implant Placement

Open the implant package.

Connect the corresponding implant mount to the implant, using the corresponding screw and tighten with the 1.25 screwdriver.

Drive the implant through the guide (using the torque ratchet) until the implant mount is engaged with the guide sleeve.

Surgical Guide Removal

After finalizing implant insertion, disconnect the mount screws and remove the implant mounts using the dedicated implant mount extractor. Remove the surgical guide.

Note

If the guide is held by lateral pins, remove the pins first. Continue with the prosthetic phase (if intended) or attach the healing caps or cover screws.

Note

*For Alpha-Bio Tec. drilling sequence, visit our website or product catalog at www.alpha-bio.net

Library support*

| Product Name | Company name | | | | |
|----------------------------------|----------------|--|--|--|--|
| Simplant | Materialise | | | | |
| Implant-Assistant | CDI | | | | |
| SiCat | Sirona | | | | |
| Ondemand3D | Cybermed | | | | |
| PointNix | PointNix | | | | |
| Implant 3D | Media Lab Srl. | | | | |
| Smart Guides | Implant Smart | | | | |
| Dental Wings | MS Dental | | | | |
| NemoBridge | Nemotec | | | | |
| Romexis | Planmenca | | | | |
| 3D Imaging Systems | Carestream | | | | |
| Anatomage Guide | Anatomage | | | | |
| Accuguide dental | Actimplant | | | | |
| Pax- i3D | Vatec | | | | |
| Implant Planinig& Guided Surgery | 3Shape | | | | |
| | Alphaimplant | | | | |

^{*}The above list is updated from time to time.



The latest updated list is available on our web site.









Note

- Implant and/or abutment libraries are available for download free of charge to every third-party technology company that would like to integrate our libraries into its systems.
- Following a short registration procedure, a full set of STL libraries and relevant metadata with detailed instructions will be made available for download from our FTP servers.
- Registered suppliers can enjoy our standard semi-annual update service to keep their systems relevant and updated at all times.



GSTK - Ordering Information*

1 SITE PREPARATION



| | Tissue Punch | | Crestal Drill | | Lateral Pin | Crestal Pin | | Ø 1.5 Milling Drill |
|----------|--------------|-------|---------------|-------|-------------|-------------|-------|---------------------|
| | small | large | small | large | | small | large | |
| Code | TPS | TPL | CDS | CDL | LP | CPS | CPL | MCD1.5 |
| Ref. no. | 65003 | 65004 | 65005 | 65006 | 65047 | 65048 | 65049 | 65050 |




2 оѕтеотому



GSTK – Ordering Information*

3 IMPLANT PLACEMENT

| | Implant Mount Internal Hex | | Implant Mount Internal Hex Screw | Implant Mount CHC | Implant Mount CHC Screw | Implant Mount Internal Hex Driver | | Implant Mount CHC Driver |
|----------|----------------------------|-------|-------------------------------------|----------------------|----------------------------|-----------------------------------|-------|-----------------------------|
| | small | | | | | small | large | |
| | | | n | | | | | |
| | 8 | В | | В | ų. | R | н | н |
| | U | U | | Ų. | | Ī | Ü | Ţ |
| | Ü | J | e e | | Ü | | | |
| Code | IMS | IML | IMHS | IMC | IMCS | IMSD | IMLD | IMCD |
| Ref. no. | 65037 | 65038 | 65039 | 65055 | 65056 | 65062 | 65063 | 65061 |



4 TOOLS & ACCESSORIES

| | Screwdriver | Hex Driver | Implant Mount Extension | Handpiece Insertion Adapter | Implant Mount Extractor | Universal Square Ratchet Head Adapter | L/S Sleeve Adapter Driver |
|----------|-------------|------------|----------------------------|--------------------------------|----------------------------|--|------------------------------|
| | short | | one | | | | |
| | | | | ñ | 9 | | |
| | 24 | # | | | | 0 | A |
| | Ψ | | A | Ä | | 5.11 | ŢŢ. |
| | N | И | U | | U | Enables use of 4 mm square driver heads | 8 |
| | | | | | | | |
| Code | HHSS1.25 | HTD1.25S | IMX | HIA | IME | USH | SAD |
| Ref. no. | 4053 | 4056 | 65042 | 65044 | 65045 | 4012 | 65057 |



Authorized European regulatory representative:

EC REP MEDES LIMITED

5 Beaumont Gate, Shenley Hill Radlett, Herts WD7 7AR, England, U.K Tel/Fax: +44 1923859810

www.alpha-bio.net

